
Midmark M11 - v1.0.0
Total Cycles: 615

Name: MIDMARK
Sterilizer ID: V2178198
01/20/2021 04:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.7 F 0.0
2:00 159.7 F 3.0
4:00 183.9 F 6.9
6:00 209.8 F 12.7
8:00 223.2 F 7.2
10:00 237.1 F 13.0
12:00 237.5 F 10.2
14:00 235.2 F 8.2
16:00 250.5 F 15.6
18:00 263.5 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.9 F 28.3
2:00 272.3 F 28.7
3:00 272.8 F 29.0
4:00 272.5 F 29.3

Min 271.3 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 228.9 F 0.0
10:00 243.7 F 0.0
15:00 229.3 F 0.0
20:00 220.5 F 0.0
25:00 220.2 F 0.0
Dry Cycle Aborted

FILLING: 1:05
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:01
DRYING: 29:53
TOTAL CYCLE: 00:57:53

01/20/2021 05:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____