

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1927

Name: RITTER2  
Sterilizer ID: V2313228  
01/12/2023 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
2:00 188.6 F 3.6  
4:00 194.3 F 8.2  
6:00 214.0 F 11.2  
8:00 222.4 F 8.8  
10:00 234.9 F 9.7  
12:00 240.1 F 12.4  
14:00 241.2 F 11.4  
16:00 254.3 F 18.4  
18:00 266.0 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.5  
1:00 272.6 F 30.3  
2:00 272.2 F 29.6  
3:00 271.5 F 29.6  
4:00 272.7 F 30.3  
  
Min 271.4 F 29.3  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.5 F 0.0  
5:00 233.1 F 0.0  
10:00 234.5 F 0.0  
15:00 227.1 F 0.0  
20:00 222.6 F 0.0  
25:00 232.3 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:01  
CYCLE COMPLETE

01/12/2023 12:18 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_