

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2356

Name: RITTER2  
Sterilizer ID: V2313228  
05/11/2023 12:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.3 F 0.0  
2:00 196.5 F 3.8  
4:00 206.5 F 9.4  
6:00 228.6 F 7.3  
8:00 233.5 F 10.9  
10:00 234.7 F 8.8  
12:00 240.0 F 9.9  
14:00 248.2 F 15.0  
16:00 260.9 F 22.0  
18:00 272.1 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.5 F 29.7  
2:00 271.4 F 29.1  
3:00 272.5 F 29.8  
4:00 271.8 F 29.2  
  
Min 271.4 F 29.0  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 205.7 F 0.0  
10:00 217.6 F 0.0  
15:00 222.8 F 0.0  
20:00 224.8 F 0.0  
25:00 225.9 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 2:31  
HEATING: 18:31  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

05/11/2023 01:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_