
Midmark M11 - v1.0.0
Total Cycles: 1873

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2022 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.7 F 0.0
2:00 129.2 F 0.0
4:00 158.2 F 2.0
6:00 181.9 F 5.2
8:00 203.4 F 9.4
10:00 220.0 F 9.3
12:00 228.1 F 7.8
14:00 237.8 F 12.7
16:00 237.3 F 9.7
18:00 237.8 F 8.9
20:00 248.6 F 14.4
22:00 260.6 F 21.1
24:00 271.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.7 F 28.9
2:00 271.7 F 28.1
3:00 272.3 F 28.7
4:00 272.5 F 28.7

Min 271.4 F 27.9
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 207.9 F 0.0
10:00 194.8 F 0.0
15:00 201.9 F 0.0
20:00 200.4 F 0.0
25:00 208.5 F 0.0
30:00 242.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:43
CYCLE COMPLETE

04/26/2022 03:26 PM
ACCEPTED: X
REJECTED:
Operator: U02