
Midmark M11 - v1.0.5
Total Cycles: 1845

Name: RITTER2
Sterilizer ID: V2313228
12/15/2022 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 195.5 F 3.3
4:00 198.0 F 8.8
6:00 224.8 F 7.4
8:00 233.2 F 10.8
10:00 234.6 F 8.7
12:00 240.2 F 9.9
14:00 249.0 F 15.1
16:00 261.7 F 22.4
18:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.3 F 29.8
2:00 271.9 F 29.0
3:00 271.9 F 29.6
4:00 272.2 F 29.4

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 212.7 F 0.0
10:00 229.9 F 0.0
15:00 232.2 F 0.0
20:00 227.5 F 0.0
25:00 225.5 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 18:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

12/15/2022 10:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____