
Midmark M11 - v1.0.5
Total Cycles: 1552

Name: RITTER2
Sterilizer ID: V2313228
10/03/2022 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.7 F 0.0
2:00 196.5 F 0.8
4:00 209.9 F 6.2
6:00 228.5 F 11.9
8:00 232.3 F 9.0
10:00 237.9 F 9.7
12:00 242.8 F 11.7
14:00 247.3 F 14.1
16:00 260.6 F 21.5
18:00 272.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.4 F 29.5
2:00 271.6 F 28.6
3:00 272.6 F 29.4
4:00 271.7 F 28.7

Min 271.3 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 233.9 F 0.0
10:00 235.9 F 0.0
15:00 227.3 F 0.0
20:00 219.5 F 0.0
25:00 216.9 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 18:21
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:17
CYCLE COMPLETE

10/03/2022 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____