
Midmark M11 - v1.0.5
Total Cycles: 239

Name: RITTER2
Sterilizer ID: V2313228
07/22/2021 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 165.0 F 2.8
4:00 179.5 F 6.7
6:00 218.9 F 11.5
8:00 223.6 F 6.0
10:00 233.3 F 11.0
12:00 233.1 F 8.2
14:00 241.7 F 11.8
16:00 245.6 F 13.4
18:00 257.6 F 19.8
20:00 268.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.3
2:00 271.3 F 28.9
3:00 272.5 F 29.5
4:00 271.6 F 28.8

Min 271.2 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 230.1 F 0.0
10:00 232.6 F 0.0
15:00 217.3 F 0.0
20:00 216.0 F 0.0
25:00 232.8 F 0.0
30:00 234.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

07/22/2021 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____