
Midmark M11 - v1.0.0
Total Cycles: 2438

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2022 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.9 F 0.0
2:00 162.5 F 2.0
4:00 166.6 F 5.5
6:00 192.6 F 9.8
8:00 212.6 F 5.1
10:00 221.0 F 7.4
12:00 233.3 F 12.0
14:00 233.1 F 8.4
16:00 241.6 F 11.4
18:00 244.1 F 12.5
20:00 256.1 F 18.8
22:00 267.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 271.7 F 29.0
2:00 272.4 F 29.2
3:00 271.4 F 28.6
4:00 272.2 F 29.4

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 213.1 F 0.0
10:00 229.1 F 0.0
15:00 228.4 F 0.0
20:00 228.8 F 0.0
25:00 229.8 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

10/17/2022 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____