
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006575
14:45 07/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:45
STERILIZING 4:42
136.4 C 220kPa 4:42
136.3 C 224kPa 5:12
136.4 C 224kPa 5:42
136.2 C 223kPa 6:12
136.4 C 222kPa 6:42
136.1 C 223kPa 7:12
136.1 C 221kPa 7:42
136.4 C 224kPa 8:12
136.4 C 224kPa 8:12
VENTING 8:12
AIR DRYING 9:06
CYCLE COMPLETE 9:10

Digital Signature #
1821AC0BBCD784A8
