
Midmark M11 - v1.0.5
Total Cycles: 3856

Name: RITTER2
Sterilizer ID: V2313228
08/14/2024 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 153.8 F 1.4
4:00 166.8 F 4.4
6:00 188.5 F 8.2
8:00 205.2 F 12.2
10:00 216.2 F 5.1
12:00 226.7 F 9.2
14:00 236.3 F 12.1
16:00 238.9 F 11.1
18:00 237.3 F 9.6
20:00 250.1 F 15.7
22:00 260.8 F 21.8
24:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.7 F 29.7
2:00 271.9 F 28.9
3:00 272.0 F 29.5
4:00 272.5 F 29.5

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 212.5 F 0.0
10:00 193.2 F 0.0
15:00 188.4 F 0.0
20:00 184.4 F 0.0
25:00 181.1 F 0.0
30:00 179.3 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:57
CYCLE COMPLETE

08/14/2024 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____