
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001913
12:17 31/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:35
STERILIZING 3:49
136.0 C 220kPa 3:49
136.4 C 226kPa 4:19
136.4 C 224kPa 4:49
136.4 C 226kPa 5:19
136.4 C 225kPa 5:49
136.1 C 223kPa 6:19
136.3 C 225kPa 6:49
136.3 C 226kPa 7:19
136.3 C 226kPa 7:19
VENTING 7:19
AIR DRYING 8:11
CYCLE COMPLETE 8:16

Digital Signature #
F9846E4ABC5C3367
