
Midmark M11 - v1.0.0
Total Cycles: 2068

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2022 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.9 F 0.0
2:00 183.1 F 3.6
4:00 221.1 F 8.4
6:00 227.2 F 8.1
8:00 232.9 F 9.6
10:00 236.3 F 8.7
12:00 244.1 F 12.6
14:00 241.9 F 11.1
16:00 255.0 F 17.7
18:00 266.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.6 F 28.9
2:00 272.0 F 28.4
3:00 271.4 F 28.5
4:00 272.8 F 29.2

Min 271.0 F 28.0
Max 273.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 230.6 F 0.0
10:00 224.6 F 0.0
15:00 224.7 F 0.0
20:00 224.2 F 0.0
25:00 225.6 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

06/28/2022 12:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____