

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3698

Name: RITTER 1  
Sterilizer ID: V2178198  
08/29/2023 12:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.0 F 0.0  
2:00 172.2 F 3.7  
4:00 199.0 F 8.2  
6:00 213.6 F 13.2  
8:00 227.6 F 7.7  
10:00 238.4 F 12.6  
12:00 237.5 F 9.8  
14:00 237.4 F 8.9  
16:00 248.7 F 14.6  
18:00 260.2 F 21.1  
20:00 271.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.2 F 28.7  
2:00 272.0 F 28.9  
3:00 271.7 F 28.3  
4:00 272.4 F 29.1  
  
Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 208.9 F 0.0  
10:00 219.8 F 0.0  
15:00 227.9 F 0.0  
20:00 227.6 F 0.0  
25:00 227.6 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 20:41  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:00:26  
CYCLE COMPLETE

08/29/2023 01:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_