
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001009
13:16 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:23
STERILIZING 3:24
136.4 C 220kPa 3:24
135.9 C 230kPa 3:54
136.2 C 232kPa 4:24
136.4 C 233kPa 4:54
136.4 C 233kPa 5:24
136.4 C 233kPa 5:54
136.4 C 233kPa 6:24
136.1 C 231kPa 6:54
136.1 C 231kPa 6:54
VENTING 6:54
AIR DRYING 7:46
CYCLE COMPLETE 8:58

Digital Signature #
9322D641793AA815
