

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3106

Name: RITTER2  
Sterilizer ID: V2313228  
11/09/2023 01:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
2:00 179.7 F 3.1  
4:00 196.4 F 8.1  
6:00 212.9 F 13.0  
8:00 227.7 F 8.1  
10:00 234.9 F 13.5  
12:00 240.5 F 11.8  
14:00 240.6 F 11.1  
16:00 254.0 F 17.9  
18:00 265.5 F 24.9  
20:00 272.6 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.5  
1:00 272.3 F 29.4  
2:00 272.9 F 29.8  
3:00 272.4 F 30.1  
4:00 271.8 F 29.4  
  
Min 271.6 F 29.3  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.4 F 0.0  
5:00 219.3 F 0.0  
10:00 210.7 F 0.0  
15:00 209.0 F 0.0  
20:00 208.5 F 0.0  
25:00 211.9 F 0.0  
30:00 215.1 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

11/09/2023 02:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_