
Midmark M11 - v1.0.0
Total Cycles: 812

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.7 F 0.0
2:00 106.7 F 0.0
4:00 121.1 F 0.0
6:00 156.3 F 0.2
8:00 181.6 F 2.2
10:00 196.0 F 5.4
12:00 209.3 F 8.7
14:00 220.9 F 12.4
16:00 226.2 F 6.9
18:00 236.5 F 11.4
20:00 234.5 F 8.5
22:00 242.0 F 11.5
24:00 245.4 F 13.0
26:00 257.2 F 19.2
28:00 267.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 272.2 F 28.5
3:00 271.7 F 28.5
4:00 272.7 F 29.1

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 205.0 F 0.0
10:00 211.3 F 0.0
15:00 211.6 F 0.0
20:00 213.3 F 0.0
25:00 214.5 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:27
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:07:46
CYCLE COMPLETE

04/27/2021 10:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____