
Midmark M11 - v1.0.5
Total Cycles: 3093

Name: RITTER2
Sterilizer ID: V2313228
11/07/2023 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.6 F 0.0
2:00 135.1 F 0.9
4:00 147.6 F 3.6
6:00 176.0 F 7.3
8:00 195.0 F 11.4
10:00 222.2 F 5.4
12:00 228.6 F 10.0
14:00 233.8 F 9.2
16:00 239.9 F 11.8
18:00 238.4 F 10.1
20:00 250.9 F 16.2
22:00 261.9 F 22.6
24:00 272.1 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.8
1:00 271.7 F 29.7
2:00 272.1 F 29.7
3:00 272.6 F 29.9
4:00 272.4 F 30.1

Min 271.3 F 29.6
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 238.7 F 0.0
10:00 226.0 F 0.0
15:00 233.2 F 0.0
20:00 223.9 F 0.0
25:00 215.3 F 0.0
30:00 180.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:16
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

11/07/2023 01:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____