
Midmark M11 - v1.0.5
Total Cycles: 994

Name: RITTER2
Sterilizer ID: V2313228
04/18/2022 06:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.5 F 0.0
2:00 74.6 F 0.0
4:00 107.4 F 0.0
6:00 155.8 F 1.1
8:00 144.9 F 4.4
10:00 168.6 F 8.1
12:00 190.0 F 11.9
14:00 211.6 F 3.2
16:00 220.4 F 6.8
18:00 231.1 F 10.8
20:00 234.9 F 7.9
22:00 241.9 F 12.4
24:00 240.2 F 10.8
26:00 252.1 F 16.8
28:00 262.9 F 23.1
30:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.7 F 29.6
2:00 271.5 F 28.7
3:00 272.4 F 29.7
4:00 272.1 F 29.1

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 197.3 F 0.0
10:00 200.7 F 0.0
15:00 201.5 F 0.0
20:00 199.6 F 0.0
25:00 200.6 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:21
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:16
CYCLE COMPLETE

04/18/2022 08:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____