

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2988

Name: RITTER 1  
Sterilizer ID: V2178198  
03/09/2023 11:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.0 F 0.0  
2:00 160.5 F 2.6  
4:00 180.2 F 6.5  
6:00 199.4 F 11.2  
8:00 218.0 F 5.6  
10:00 229.5 F 10.6  
12:00 230.2 F 7.0  
14:00 243.5 F 12.9  
16:00 241.8 F 11.3  
18:00 254.9 F 17.8  
20:00 266.1 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.6 F 29.0  
2:00 271.4 F 28.2  
3:00 272.2 F 29.0  
4:00 272.2 F 28.7  
  
Min 271.1 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 224.7 F 0.0  
10:00 234.8 F 0.0  
15:00 231.7 F 0.0  
20:00 225.0 F 0.0  
25:00 230.0 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 21:40  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:47  
CYCLE COMPLETE

03/09/2023 12:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_