
Midmark M11 - v1.0.0
Total Cycles: 5317

Name: RITTER 1
Sterilizer ID: V2178198
02/05/2025 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.6 F 0.0
2:00 91.3 F 0.8
4:00 110.6 F 2.8
6:00 142.8 F 5.0
8:00 166.6 F 7.7
10:00 185.1 F 10.8
12:00 209.0 F 5.5
14:00 212.8 F 5.9
16:00 224.3 F 9.9
18:00 235.6 F 9.1
20:00 235.8 F 9.5
22:00 241.0 F 10.4
24:00 243.2 F 12.1
26:00 254.4 F 17.7
28:00 264.1 F 23.5
30:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.1 F 28.7
2:00 272.2 F 29.2
3:00 271.5 F 28.5
4:00 272.1 F 28.9

Min 271.4 F 28.4
Max 272.3 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 199.8 F 0.0
10:00 213.0 F 0.0
15:00 203.1 F 0.0
20:00 198.6 F 0.0
25:00 204.0 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:08:46
CYCLE COMPLETE

02/05/2025 03:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____