
Midmark M11 - v1.0.5
Total Cycles: 2245

Name: RITTER2
Sterilizer ID: V2313228
04/13/2023 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 184.9 F 3.1
4:00 198.6 F 8.3
6:00 217.8 F 10.4
8:00 229.8 F 9.9
10:00 233.3 F 8.5
12:00 242.3 F 12.1
14:00 244.8 F 13.3
16:00 257.7 F 20.1
18:00 268.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.8
2:00 271.9 F 29.1
3:00 271.6 F 29.4
4:00 272.6 F 29.8

Min 271.3 F 28.8
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 219.4 F 0.0
10:00 227.4 F 0.0
15:00 228.0 F 0.0
20:00 228.0 F 0.0
25:00 228.7 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

04/13/2023 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____