
Midmark M11 - v1.0.0
Total Cycles: 4998

Name: RITTER 1
Sterilizer ID: V2178198
10/16/2024 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.3 F 0.0
2:00 137.9 F 0.0
4:00 140.9 F 0.1
6:00 178.9 F 0.5
8:00 202.4 F 3.3
10:00 218.5 F 7.0
12:00 226.0 F 10.7
14:00 232.9 F 7.2
16:00 233.5 F 10.2
18:00 239.0 F 8.8
20:00 244.3 F 13.1
22:00 243.2 F 11.8
24:00 255.1 F 18.1
26:00 265.9 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.8 F 29.1
2:00 271.9 F 28.4
3:00 272.4 F 29.1
4:00 272.4 F 28.9

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 207.7 F 0.0
10:00 197.3 F 0.0
15:00 190.8 F 0.0
20:00 234.2 F 0.0
25:00 214.8 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 27:46
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:06:22
CYCLE COMPLETE

10/16/2024 11:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____