
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003039
11:44 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:10
STERILIZING 4:10
136.5 C 222kPa 4:10
136.4 C 223kPa 4:40
136.1 C 220kPa 5:10
136.2 C 221kPa 5:40
136.3 C 224kPa 6:10
136.3 C 223kPa 6:40
136.4 C 224kPa 7:10
CYCLE ABORTED 7:29
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:29

Digital Signature #
9BA3DAF70A20F3CB
