
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001793
13:44 13/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:34
STERILIZING 3:39
136.0 C 220kPa 3:39
136.4 C 222kPa 4:09
136.5 C 226kPa 4:39
136.4 C 223kPa 5:09
136.0 C 222kPa 5:39
136.4 C 224kPa 6:09
136.3 C 223kPa 6:39
136.3 C 224kPa 7:09
136.3 C 224kPa 7:09
VENTING 7:09
AIR DRYING 7:58
CYCLE COMPLETE 37:59

Digital Signature #
48FC19918C4ABFD3
