
Midmark M11 - v1.0.5
Total Cycles: 4251

Name: RITTER2
Sterilizer ID: V2313228
01/09/2025 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.6 F 0.0
2:00 108.5 F 0.0
4:00 109.6 F 0.0
6:00 146.7 F 0.5
8:00 165.2 F 1.6
10:00 191.1 F 4.9
12:00 212.4 F 8.4
14:00 219.3 F 12.4
16:00 226.3 F 7.6
18:00 236.4 F 12.3
20:00 236.3 F 9.8
22:00 238.1 F 9.8
24:00 246.9 F 14.0
26:00 257.8 F 20.0
28:00 267.7 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.4 F 29.3
2:00 271.5 F 29.2
3:00 272.1 F 29.1
4:00 271.7 F 29.4

Min 271.1 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 213.6 F 0.0
10:00 191.0 F 0.0
15:00 190.8 F 0.0
20:00 208.6 F 0.0
25:00 222.7 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:25
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:08:02
CYCLE COMPLETE

01/09/2025 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____