
Midmark M11 - v1.0.0
Total Cycles: 2452

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2022 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.9 F 0.0
2:00 201.2 F 4.1
4:00 224.2 F 10.2
6:00 227.0 F 5.7
8:00 239.1 F 12.2
10:00 237.8 F 10.3
12:00 237.5 F 9.0
14:00 251.8 F 16.4
16:00 264.4 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.1
2:00 271.5 F 28.3
3:00 272.6 F 29.3
4:00 272.2 F 28.8

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 236.2 F 0.0
10:00 231.1 F 0.0
15:00 230.5 F 0.0
20:00 228.9 F 0.0
25:00 228.4 F 0.0
30:00 208.6 F 0.0
Drying Complete

FILLING: 3:54
HEATING: 17:54
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

10/18/2022 03:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____