
Midmark M11 - v1.0.5
Total Cycles: 169

Name: RITTER2
Sterilizer ID: V2313228
07/06/2021 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.6 F 0.0
2:00 189.6 F 3.2
4:00 200.8 F 8.2
6:00 218.3 F 13.0
8:00 227.3 F 9.2
10:00 235.9 F 10.2
12:00 242.6 F 12.8
14:00 242.6 F 11.9
16:00 254.8 F 18.3
18:00 265.7 F 24.9
20:00 272.3 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.6 F 29.6
2:00 272.8 F 30.0
3:00 271.9 F 29.3
4:00 273.0 F 29.9

Min 271.8 F 29.2
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 226.8 F 0.0
10:00 222.0 F 0.0
15:00 224.1 F 0.0
20:00 224.9 F 0.0
25:00 226.6 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

07/06/2021 03:36 PM
ACCEPTED: X
REJECTED:
Operator: U02