

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007290  
7:22 30/04/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:13  
STERILIZING 4:28  
136.4 C 220kPa 4:28  
136.3 C 222kPa 4:58  
136.4 C 219kPa 5:28  
136.1 C 222kPa 5:58  
136.1 C 222kPa 6:28  
136.4 C 222kPa 6:58  
136.4 C 223kPa 7:28  
136.4 C 223kPa 7:58  
136.4 C 223kPa 7:58  
VENTING 7:58  
AIR DRYING 8:52  
CYCLE COMPLETE 9:10  
-----

Digital Signature #  
840A764E41A31D12  
-----