
Midmark M11 - v1.0.0
Total Cycles: 3842

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2023 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.4 F 0.0
2:00 179.5 F 0.1
4:00 197.6 F 3.8
6:00 211.5 F 8.3
8:00 225.4 F 13.2
10:00 233.0 F 9.1
12:00 238.4 F 10.3
14:00 243.7 F 12.5
16:00 242.4 F 11.3
18:00 255.1 F 17.8
20:00 266.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.6
2:00 271.8 F 28.5
3:00 272.0 F 28.4
4:00 271.8 F 28.7

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 221.5 F 0.0
10:00 215.8 F 0.0
15:00 215.3 F 0.0
20:00 213.9 F 0.0
25:00 213.6 F 0.0
30:00 213.3 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

09/29/2023 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____