
Midmark M11 - v1.0.0
Total Cycles: 3700

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2023 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 188.2 F 3.9
4:00 204.6 F 8.8
6:00 223.5 F 8.5
8:00 229.7 F 9.1
10:00 237.5 F 10.4
12:00 242.0 F 12.0
14:00 240.4 F 10.7
16:00 253.5 F 17.2
18:00 265.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 271.9 F 28.4
2:00 272.5 F 28.9
3:00 271.9 F 28.8
4:00 271.7 F 28.4

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 211.3 F 0.0
10:00 220.9 F 0.0
15:00 229.2 F 0.0
20:00 234.5 F 0.0
25:00 232.5 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 2:59
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

08/29/2023 03:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____