
Midmark M11 - v1.0.5
Total Cycles: 182

Name: RITTER2
Sterilizer ID: V2313228
07/08/2021 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.2 F 0.0
2:00 129.1 F 1.9
4:00 155.4 F 5.7
6:00 186.5 F 10.8
8:00 208.6 F 4.3
10:00 224.6 F 10.0
12:00 231.2 F 7.9
14:00 242.3 F 13.5
16:00 243.1 F 12.5
18:00 256.6 F 19.5
20:00 268.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.6 F 29.7
2:00 272.4 F 29.9
3:00 272.2 F 29.3
4:00 273.0 F 30.1

Min 271.7 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 233.4 F 0.0
10:00 225.0 F 0.0
15:00 221.2 F 0.0
20:00 217.5 F 0.0
25:00 216.1 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 21:14
STERILIZING: 4:00
VENTING: 2:51
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

07/08/2021 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____