
Midmark M11 - v1.0.5
Total Cycles: 4027

Name: RITTER2
Sterilizer ID: V2313228
10/15/2024 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 70.1 F 0.0
4:00 79.5 F 0.0
6:00 107.7 F 0.1
8:00 144.6 F 0.4
10:00 173.0 F 1.8
12:00 195.5 F 5.0
14:00 203.1 F 8.9
16:00 218.8 F 12.8
18:00 225.5 F 7.4
20:00 236.3 F 12.0
22:00 235.2 F 9.3
24:00 239.3 F 10.0
26:00 247.0 F 14.1
28:00 258.4 F 20.4
30:00 269.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.7 F 29.4
2:00 271.5 F 29.0
3:00 272.5 F 29.4
4:00 271.4 F 29.0

Min 271.3 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 211.2 F 0.0
10:00 190.2 F 0.0
15:00 190.5 F 0.0
20:00 187.5 F 0.0
25:00 186.4 F 0.0
30:00 185.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:23
CYCLE COMPLETE

10/15/2024 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____