
Midmark M11 - v1.0.5
Total Cycles: 2312

Name: RITTER2
Sterilizer ID: V2313228
04/28/2023 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.2 F 0.0
2:00 183.3 F 2.9
4:00 203.8 F 8.3
6:00 220.5 F 13.5
8:00 230.0 F 9.0
10:00 237.2 F 10.4
12:00 243.5 F 12.9
14:00 243.6 F 12.3
16:00 256.3 F 19.0
18:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.4 F 29.6
2:00 271.6 F 28.7
3:00 272.4 F 29.6
4:00 271.9 F 29.0

Min 271.3 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 216.0 F 0.0
10:00 205.7 F 0.0
15:00 199.9 F 0.0
20:00 196.6 F 0.0
25:00 194.3 F 0.0
30:00 192.4 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

04/28/2023 10:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____