

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2733

Name: RITTER 1  
Sterilizer ID: V2178198  
01/04/2023 11:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
2:00 198.1 F 0.0  
4:00 207.2 F 3.9  
6:00 227.1 F 9.1  
8:00 233.3 F 8.9  
10:00 237.6 F 11.1  
12:00 236.1 F 8.9  
14:00 239.7 F 9.5  
16:00 249.1 F 14.7  
18:00 261.6 F 21.8  
20:00 272.8 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 28.9  
2:00 271.4 F 28.2  
3:00 272.7 F 29.0  
4:00 271.8 F 28.3  
  
Min 271.3 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.5 F 0.0  
5:00 225.5 F 0.0  
10:00 247.9 F 0.0  
15:00 242.3 F 0.0  
20:00 242.6 F 0.0  
25:00 229.3 F 0.0  
30:00 221.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 20:23  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:23  
CYCLE COMPLETE

01/04/2023 12:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_