
Midmark M11 - v1.0.5
Total Cycles: 1631

Name: RITTER2
Sterilizer ID: V2313228
10/20/2022 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.7 F 0.0
2:00 174.4 F 3.1
4:00 189.9 F 8.0
6:00 207.5 F 10.8
8:00 224.9 F 9.1
10:00 235.3 F 9.6
12:00 241.6 F 13.0
14:00 242.1 F 12.2
16:00 255.4 F 19.1
18:00 267.0 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.8 F 30.2
2:00 272.3 F 29.6
3:00 271.7 F 29.6
4:00 273.0 F 30.2

Min 271.6 F 29.2
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 204.7 F 0.0
10:00 215.2 F 0.0
15:00 221.4 F 0.0
20:00 220.9 F 0.0
25:00 219.4 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 19:28
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

10/20/2022 12:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____