
Midmark M11 - v1.0.0
Total Cycles: 4337

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2024 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.6 F 0.0
2:00 154.1 F 1.8
4:00 155.5 F 4.7
6:00 176.9 F 8.2
8:00 196.7 F 12.2
10:00 211.2 F 4.5
12:00 221.0 F 8.7
14:00 232.6 F 13.2
16:00 235.0 F 9.4
18:00 239.6 F 10.4
20:00 245.1 F 13.0
22:00 256.6 F 19.0
24:00 266.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.4 F 28.3
3:00 272.3 F 29.2
4:00 271.9 F 28.7

Min 271.2 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 213.1 F 0.0
10:00 219.9 F 0.0
15:00 224.0 F 0.0
20:00 217.0 F 0.0
25:00 230.7 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:37
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:04:12
CYCLE COMPLETE

02/27/2024 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____