

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001744  
12:58 17/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:07  
STERILIZING 3:08  
136.0 C 220kPa 3:08  
136.3 C 226kPa 3:38  
136.4 C 226kPa 4:08  
136.3 C 226kPa 4:38  
136.4 C 222kPa 5:08  
136.4 C 224kPa 5:38  
136.4 C 224kPa 6:08  
136.1 C 223kPa 6:38  
136.1 C 223kPa 6:38  
VENTING 6:38  
AIR DRYING 7:27  
CYCLE COMPLETE 11:10  
-----

Digital Signature #  
D78483D48A70BAF2  
-----