
Midmark M11 - v1.0.5
Total Cycles: 3860

Name: RITTER2
Sterilizer ID: V2313228
08/14/2024 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.4 F 0.0
2:00 170.8 F 3.2
4:00 190.1 F 7.6
6:00 205.7 F 12.7
8:00 225.5 F 7.6
10:00 233.1 F 12.8
12:00 237.5 F 10.6
14:00 236.6 F 9.3
16:00 250.3 F 16.1
18:00 261.9 F 22.7
20:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.3 F 29.6
2:00 272.1 F 30.0
3:00 271.6 F 29.7
4:00 272.3 F 29.6

Min 271.5 F 29.4
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 221.9 F 0.0
10:00 220.9 F 0.0
15:00 220.9 F 0.0
20:00 202.0 F 0.0
25:00 197.1 F 0.0
30:00 199.1 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 21:11
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

08/14/2024 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____