

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1993

Name: RITTER2  
Sterilizer ID: V2313228  
02/01/2023 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
2:00 177.4 F 3.6  
4:00 187.4 F 8.1  
6:00 207.3 F 13.5  
8:00 220.8 F 8.0  
10:00 234.9 F 13.6  
12:00 239.0 F 11.2  
14:00 238.5 F 10.1  
16:00 252.6 F 17.1  
18:00 264.6 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.3 F 29.5  
2:00 271.5 F 29.0  
3:00 272.4 F 29.5  
4:00 271.6 F 28.8  
  
Min 271.4 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 218.0 F 0.0  
10:00 227.9 F 0.0  
15:00 228.6 F 0.0  
20:00 225.2 F 0.0  
25:00 226.9 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 19:52  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:27  
CYCLE COMPLETE

02/01/2023 12:58 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_