
Midmark M11 - v1.0.5
Total Cycles: 4205

Name: RITTER2
Sterilizer ID: V2313228
12/11/2024 04:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.9 F 0.0
2:00 96.4 F 0.0
4:00 123.5 F 0.6
6:00 157.6 F 3.7
8:00 177.5 F 7.3
10:00 196.2 F 11.1
12:00 211.0 F 4.9
14:00 220.8 F 7.4
16:00 231.8 F 11.5
18:00 231.4 F 8.3
20:00 243.1 F 12.6
22:00 243.1 F 12.6
24:00 255.0 F 18.8
26:00 266.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 273.0 F 30.0
2:00 272.0 F 29.2
3:00 272.7 F 30.2
4:00 272.6 F 29.7

Min 271.7 F 29.1
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 212.0 F 0.0
10:00 220.0 F 0.0
15:00 217.6 F 0.0
20:00 214.2 F 0.0
25:00 213.2 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 27:41
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:06:19
CYCLE COMPLETE

12/11/2024 05:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____