
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002635
8:24 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:09
STERILIZING 3:17
136.3 C 222kPa 3:17
136.3 C 223kPa 3:47
136.4 C 225kPa 4:17
136.4 C 225kPa 4:47
136.4 C 223kPa 5:17
136.0 C 223kPa 5:47
136.3 C 222kPa 6:17
136.4 C 225kPa 6:47
136.4 C 225kPa 6:47
VENTING 6:47
AIR DRYING 7:38
CYCLE COMPLETE 14:30

Digital Signature #
73DA8033A681EFB8
