
Midmark M11 - v1.0.5
Total Cycles: 4479

Name: RITTER2
Sterilizer ID: V2313228
04/02/2025 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.3 F 0.0
2:00 143.4 F 1.6
4:00 144.9 F 4.3
6:00 162.4 F 7.9
8:00 186.1 F 12.0
10:00 208.2 F 4.8
12:00 221.6 F 9.0
14:00 231.7 F 12.2
16:00 237.4 F 10.9
18:00 235.5 F 9.1
20:00 248.8 F 15.5
22:00 260.5 F 21.9
24:00 270.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.5
1:00 272.0 F 29.7
2:00 271.6 F 30.0
3:00 271.4 F 29.9
4:00 271.2 F 29.4

Min 271.1 F 29.4
Max 272.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 207.6 F 0.0
10:00 203.1 F 0.0
15:00 190.4 F 0.0
20:00 194.4 F 0.0
25:00 193.8 F 0.0
30:00 191.9 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:24
CYCLE COMPLETE

04/02/2025 03:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____