
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001044
10:37 21/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:37
STERILIZING 5:06
136.4 C 222kPa 5:06
136.3 C 224kPa 5:36
136.3 C 223kPa 6:06
136.2 C 222kPa 6:36
136.4 C 225kPa 7:06
136.4 C 225kPa 7:36
136.3 C 223kPa 8:06
136.2 C 222kPa 8:36
136.2 C 222kPa 8:36
VENTING 8:36
AIR DRYING 9:31
CYCLE COMPLETE 12:22

Digital Signature #
4091E7CD615383D3
