

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000767  
11:06 28/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:13  
STERILIZING 4:31  
136.5 C 233kPa 4:31  
136.3 C 233kPa 5:01  
136.3 C 233kPa 5:31  
136.3 C 232kPa 6:01  
136.2 C 231kPa 6:31  
136.4 C 232kPa 7:01  
136.1 C 230kPa 7:31  
136.4 C 233kPa 8:01  
136.4 C 233kPa 8:01  
VENTING 8:01  
AIR DRYING 8:58  
CYCLE COMPLETE 8:59  
-----

Digital Signature #  
252BED440B7D3E46  
-----