
Midmark M11 - v1.0.5
Total Cycles: 2262

Name: RITTER2
Sterilizer ID: V2313228
04/18/2023 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 171.6 F 3.6
4:00 181.7 F 7.7
6:00 208.7 F 13.0
8:00 223.4 F 8.2
10:00 236.4 F 12.3
12:00 238.8 F 11.5
14:00 238.4 F 10.4
16:00 252.3 F 17.2
18:00 264.0 F 24.2
20:00 271.3 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.6
1:00 272.7 F 30.4
2:00 272.2 F 29.7
3:00 272.7 F 30.4
4:00 272.4 F 29.9

Min 271.7 F 29.5
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 214.3 F 0.0
10:00 225.8 F 0.0
15:00 226.4 F 0.0
20:00 225.7 F 0.0
25:00 227.3 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

04/18/2023 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____