
Midmark M11 - v1.0.5
Total Cycles: 4211

Name: RITTER2
Sterilizer ID: V2313228
12/13/2024 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.5 F 0.0
2:00 147.2 F 1.2
4:00 166.5 F 3.6
6:00 179.4 F 7.3
8:00 194.2 F 11.0
10:00 218.6 F 4.4
12:00 225.7 F 8.0
14:00 234.9 F 12.4
16:00 236.3 F 9.3
18:00 240.9 F 10.6
20:00 245.7 F 13.3
22:00 256.9 F 19.2
24:00 267.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.1
2:00 271.4 F 29.0
3:00 272.1 F 29.0
4:00 271.5 F 29.0

Min 271.1 F 28.6
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 247.4 F 0.0
10:00 194.1 F 0.0
15:00 183.6 F 0.0
20:00 178.7 F 0.0
25:00 176.6 F 0.0
30:00 175.4 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 25:36
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:04:40
CYCLE COMPLETE

12/13/2024 12:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____