
Midmark M11 - v1.0.0
Total Cycles: 2768

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2023 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.6 F 0.0
2:00 178.4 F 0.0
4:00 199.1 F 4.8
6:00 217.2 F 9.8
8:00 227.2 F 7.1
10:00 237.3 F 11.1
12:00 235.5 F 8.7
14:00 239.8 F 9.8
16:00 248.6 F 14.3
18:00 260.7 F 21.1
20:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.9
2:00 271.8 F 28.1
3:00 272.1 F 28.8
4:00 272.5 F 28.7

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 225.2 F 0.0
10:00 244.5 F 0.0
15:00 239.7 F 0.0
20:00 227.4 F 0.0
25:00 222.7 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

01/11/2023 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____