

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1429

Name: RITTER 1  
Sterilizer ID: V2178198  
12/07/2021 12:10 PM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.1 F 0.0  
2:00 137.6 F 0.7  
4:00 164.5 F 3.6  
6:00 190.8 F 8.1  
8:00 207.3 F 9.1  
10:00 227.9 F 8.3  
12:00 237.6 F 10.5  
14:00 243.0 F 12.1  
16:00 240.9 F 10.5  
18:00 254.7 F 17.5  
20:00 266.5 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.2 F 28.2  
2:00 272.8 F 28.7  
3:00 272.0 F 28.3  
  
Min 271.5 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 227.2 F 0.0  
5:00 177.6 F 0.0  
10:00 154.9 F 0.0  
15:00 142.9 F 0.0  
20:00 132.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:26  
HEATING: 21:28  
STERILIZING: 3:00  
VENTING: 2:52  
DRYING: 22:37  
TOTAL CYCLE: 00:51:23

12/08/2021 07:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U08