

-----  
Midmark M11 - v1.0.5  
Total Cycles: 883

Name: RITTER2  
Sterilizer ID: V2313228  
03/16/2022 11:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.2 F 0.0  
2:00 121.8 F 1.2  
4:00 135.4 F 3.6  
6:00 162.0 F 7.1  
8:00 190.7 F 11.2  
10:00 207.0 F 4.9  
12:00 230.9 F 9.4  
14:00 233.3 F 10.4  
16:00 237.1 F 11.0  
18:00 235.7 F 9.1  
20:00 248.1 F 15.4  
22:00 259.3 F 21.7  
24:00 270.1 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.5  
1:00 271.9 F 30.3  
2:00 272.9 F 30.1  
3:00 271.7 F 29.6  
4:00 271.0 F 29.4

Min 271.0 F 29.3  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 223.0 F 0.0  
10:00 233.2 F 0.0  
15:00 229.1 F 0.0  
20:00 228.2 F 0.0  
25:00 231.1 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:59  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:03:19  
CYCLE COMPLETE

03/16/2022 12:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_