

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1704

Name: RITTER2  
Sterilizer ID: V2313228  
11/09/2022 10:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
2:00 188.0 F 2.8  
4:00 200.4 F 7.8  
6:00 218.9 F 13.5  
8:00 230.8 F 9.5  
10:00 234.6 F 8.8  
12:00 243.4 F 12.2  
14:00 246.0 F 13.4  
16:00 258.9 F 20.5  
18:00 270.5 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.6 F 29.3  
2:00 271.5 F 28.6  
3:00 272.0 F 29.3  
4:00 272.6 F 29.3  
  
Min 271.1 F 28.5  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 219.9 F 0.0  
10:00 232.0 F 0.0  
15:00 230.1 F 0.0  
20:00 227.5 F 0.0  
25:00 226.3 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 18:52  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

11/09/2022 11:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_